


**Wellbrook ALA1530**  
Typical Calibration Data



**The above graph shows the Antenna Factor K; V/m to V out at 50 Ohms.**

**Wellbrook ALA1530 Typical Calibration Data**

Frequency (MHz)	Correction for dBuV/m (dB)
0.01	66.6
0.03	44.2
0.05	33.6
0.08	25.9
0.1	22.9
0.12	20.8
0.15	18.5
0.18	16.9
0.19	16.5
0.21	15.7
0.25	14.4
0.3	13.3
0.35	12.5
0.4	12
0.45	11.5
0.5	11.1
0.6	10.8
0.8	10.2
1	9.8
1.5	9.6
2	9.2
3	9.2
4	8.4
5	8.2
6	8.2
8	7.7
10	7.6
12	7.3
14	7.2
16	7.1
18	7
20	7.2
22	7.3
24	7.4
26	7.6
28	7.8
30	7.8